С Ν С М E М E Т G

Photobiomodulation Therapy for Wound Care: A Potent, Noninvasive, Photoceutical Approach





1.5 Contact Hours

Rodrigo Crespo Mosca, PhD, DDS, MSc • Postdoctoral Associate • Oral Biology & Biomedical Engineering • School of Dental Medicine, Engineering & Applied Sciences • University of Buffalo • Buffalo, New York

Adrian A. Ong, MD • Resident Physician • Department of Otolaryngology • University of Buffalo • Buffalo, New York

Omar Albasha • Dental Student • Oral Biology & Biomedical Engineering • University of Buffalo • Buffalo, New York

Kathryn Bass, MD, MBA • Attending Pediatric Surgeon • Pediatric Surgery • Oshei Children's Hospital • University of Buffalo • Buffalo, New York

Praveen Arany, PhD, BDS, MDS, MMSc • Assistant Professor • Oral Biology & Biomedical Engineering • School of Dental Medicine, Engineering & Applied Sciences • University of Buffalo • Buffalo, New York

The author, faculty, staff, and planners, including spouses/partners (if any), in any position to control the content of this CME/CNE activity have disclosed that they have no financial relationships with, or financial interests in, any commercial companies relevant to this educational activity.

To earn CME credit, you must read the CME article and complete the quiz online, answering at least 13 of the 18 questions correctly.

This continuing educational activity will expire for physicians on March 31, 2021, and for nurses on March 5, 2021.

All tests are now online only; take the test at http://cme.lww.com for physicians and www.nursingcenter.com for nurses. Complete CE/CME information is on the last page of this article. **GENERAL PURPOSE:**

To provide background and examine evidence for the therapeutic application of light energy treatments for wound healing.

TARGET AUDIENCE:

This continuing education activity is intended for physicians, physician assistants, nurse practitioners, and nurses with an interest in skin and wound care.

LEARNING OBJECTIVES/OUTCOMES:

After completing this continuing education activity, you should be better able to:

1. Explain the basics of lasers, light-emitting diodes, and light-tissue interactions as they apply to photobiomodulation therapy.

2. Summarize the results of the authors' literature review of the evidence regarding the therapeutic applications of photobiomodulation treatments for wound healing.

ABSTRACT

OBJECTIVE: To provide background and examine evidence for the therapeutic applications of light energy treatments for wound healing.

METHODS: A search was performed in PubMed for peer-reviewed scientific articles published in the last 5 years using the search terms "photobiomodulation therapy" and "low-level laser therapy," and these terms combined with "wound," using a "human species" filter. This search yielded 218 articles on photobiomodulation therapy or low-level laser therapy and wounds. Of these, only articles on in vivo wound care using light treatments were specifically included in this review (n = 11).

RESULTS: The wound healing effects of low-dose laser treatments were first described over 50 years ago. Various doses ranging from 0.1 to 10 J/cm² and wavelengths ranging from 405 to 1,000 nm appear to provide therapeutic benefits for a broad range of chronic wounds. A range of light energy sources from LEDs to lasers have been used and have specific advantages and limitations. There is a lack of consensus on standardized treatment parameters such as wavelengths, dose, and therapeutic outcomes in the reviewed studies, preventing direct comparison and clinical protocol recommendation. An expert opinion based on ongoing research studies and reported literature is offered.

CONCLUSIONS: Noninvasive, economical, and multipurpose light devices are an attractive tool for wound management. However, there is an urgent need in the wound care community to develop optimal clinical protocols for use based on well-designed, rigorous clinical research studies.

KEYWORDS: LASER, LED, literature review, LLLT, low-level laser therapy, PBM, photobiomodulation therapy, wound healing

ADV SKIN WOUND CARE 2019;32:157-67.

INTRODUCTION

The use of light therapy dates back to ancient civilizations, going as far back as the ancient Egyptians and Indians, who used sunlight (heliotherapy) for healing and promoting health.¹ The therapeutic use of light energy was more fully appreciated in the late 19th century when a Danish physician-scientist, Niels Ryberg Finsen, demonstrated the benefits of red and blue light in the treatment of lupus vulgaris and was recognized with the 1903 Nobel Prize in Medicine and Physiology.² In 1960, the L.A.S.E.R. (Light Amplification by Stimulated Emission of Radiation) by Theodore Maiman was invented, based on theoretical work by Albert Einstein in 1917. This brought renewed attention to the therapeutic light energy field.^{3,4} The monochromatic, coherent, and collimated nature of lasers led to immediate interest in their biologic effects. In 1967, Endre Mester,^{5,6} a Hungarian physicianscientist, reported that low-dose laser treatments were capable of promoting wound healing and hair regrowth in mice. He termed this phenomenon *photostimulation* and went on to demonstrate the efficacy of this treatment in human patients with skin ulcers.⁷

EVOLUTION AND CURRENT CONSENSUS ON TREATMENT TERMINOLOGY

Several terms have been used to describe low-dose light treatments. The most popular include cold laser therapy and lowlevel laser therapy (LLLT). However, these are poor descriptors; there is no actual "cooling" during these treatments, and the terms "low" and "level" are vague. Moreover, there is considerable evidence for the utility of nonlaser devices in mediating these therapeutic benefits, indicating that the term "laser" is inaccurate. Hence, in 2014, the North American Association for Light Therapy and the World Association for Laser Therapy agreed by consensus to promote the term *photobiomodulation* (PBM) therapy as

a form of light treatment that utilizes nonionizing forms of light sources, including lasers, light emitting diodes (LEDs) and broadband light, in the visible and infrared spectrum, involving a nonthermal process with endogenous chromophores eliciting photophysical (i.e., linear and nonlinear) and photochemical events at various biological scales. This treatment results in beneficial therapeutic outcomes including, but not limited to, the alleviation of pain or inflammation, immunomodulation, and promotion of wound healing and tissue regeneration.⁸

DEVICES FOR PBM THERAPY

Significant improvements in laser technologies have enabled a dramatic reduction in footprint (size) and cost of laser technology since its invention in the 1960s. The early gas (helium-neon) or crystal (ruby, alexandrite) lasers have been largely replaced with newer, compact, and more economical semiconductor-based lasers (eg, gallium-aluminum-arsenide, aluminum gallium indium phosphide, indium gallium arsenide, and gallium-arsenide, among others).9 This new semiconductor technology has also enabled the development of several newer wavelengths, spanning a broad range from ultraviolet to the far-infrared spectrum (Figure 1). In the 1990s, the introduction of solid-state, high-efficiency LEDs revolutionized the lighting industry, including biomedical applications of light.¹⁰ The use of LEDs for PBM therapy has become more widespread, and its clinical efficacy has been clearly demonstrated.¹¹ However, there are several questions about the physical nature of LEDs and the precise biologic responses they elicit compared with lasers (Table 1). The clinical implications of coherence, polarization, half-width full-maximum, beam divergence, and

ADVANCES IN SKIN & WOUND CARE • VOL. 32 NO. 4 **158** Copyright © 2019 Wolters Kluwer Health, Inc. All rights reserved.





the potential thermal interactions of various PBM devices on clinical safety and efficacy remain to be fully explored.^{12–15}

LIGHT-TISSUE INTERACTIONS: A PHOTOCEUTICAL APPROACH TO **PBM THERAPY**

Light is a physical form of energy. To understand its therapeutic applications, a fundamental understanding of its biologic interactions is pertinent. A brief overview of light therapy terminology is presented (Table 2), and the reader is referred to more detailed descriptions of the physical phenomenon and characteristics of biologic tissue interactions.^{11,16,17}

Whereas the high-power effects of lasers are well characterized, the low-power applications are still being investigated. Certain areas have benefitted from from fundamental explorations of visual phototransduction, sunlight and vitamin D metabolism, and melatonin and pineal gland biology of circadian rhythms. Various applications are being actively investigated, and some

Table 1.

COMPARISON OF LASER AND LED DEVICES USED FOR PBM THERAPT	COMPARISON	OF LASER	AND LED	DEVICES	USED	FOR PBM	THERAPY
--	------------	----------	---------	---------	------	---------	---------

	LASER
Electroluminance	Stimulated emission
Light-emitting diode	Light amplification by stimulated emission of radiation
Slower response	Faster response
50–100 mA	5–40 mA
Noncoherent, usually noncollimated, includes color shades (larger HWFM); nonpolarized	Coherent, collimated, monochromatic (very small HWFM); polarized
Wide junction area	Narrow and small junction
10–50 THz	1–2 MHz
Approximately 20%	Approximately 70%
High	Extremely low
Low cost, broader applications	High cost, specific application
	lectroluminance ght-emitting diode lower response 0–100 mA oncoherent, usually noncollimated, includes olor shades (larger HWFM); nonpolarized /ide junction area 0–50 THz pproximately 20% ligh ow cost, broader applications

Abbreviations: HWFM, half width of full maximum; PBM, photobiomodulation

ADVANCES IN SKIN & WOUND CARE · APRIL 2019 Copyright © 2019 Wolters Kluwer Health, Inc. All rights reserved.

Table 2. BASIC TERMINOLOGY FOR PBM TREATMENTS AND LIGHT-TISSUE INTERACTIONS

Parameter	Unit	Description
Wavelength	nm	Wavelength is a measurement of how far a photon will travel in one complete cycle; this can vary from ultrashort cosmic rays (femtometers) to radiofrequency (kilometers). Visible and infrared light are usually referred to in nanometers.
Energy	J	Each photon carries a discrete amount (quantum) of energy (E), where $E = h V$ ($E = energy$, $h = Planck$ constant, $V =$ velocity). Therefore, a short wavelength = high-frequency and high-energy photons; a long wavelength = low-frequency and low-energy photons.
Fluence	J/cm ²	Refers to energy density, calculated as energy (J) delivered per unit area (cm ²)
Power	W	Power is the rate of delivery of the energy, $1 \text{ W} = 1 \text{ J/s}$
Irradiance	W/cm ²	Refers to power density, calculated as power (W) over unit area (cm ²)
Continuous wave (CW)	_	Energy is transferred in a continuous mode, beam always "on"
Duty cycle	%	Amount of time light source is actually "on." A CW laser is always "on" 100% duty cycle
Pulsing	Peak power (W), pulse frequency (Hz), pulse width (s), pulse interval (s) and pulse train (s/min)	The energy transfer from a beam can be chopped or divided, and several parameters need to be reported: peak power is stated in watts; frequency is reported in hertz, which refers to number of pulses/s; a pulse width is duration the beam is turned on in seconds; a pulse interval is time pulse is off (s) and a pulse train is a series of pulse widths and gaps during single event (treatment). Average power (W) x of a pulsed beam is calculated as peak power (W) pulse width (s) x pulse frequency (Hz); this is especially important in dispersing thermal energy during treatment and preventing inadvertent damage.
Frequency	Hertz (Hz)	Refers to cycles/s, inversely related to wavelength as frequency (γ) = speed of light (c)/wavelength (λ). Therefore, a short wavelength = high frequency and a long wavelength = low frequency.
Chromophore	—	The part of the molecule that is responsible for its color; lasers are monochromatic (single color; peak wavelength).
Photon	_	An elementary particle of electromagnetic radiation that has no mass and can transfer energy.
Collimation	_	All beams are directed in a similar direction, and the overall beam diameter is small.
Coherence	_	Photons are vibrating in similar phase (space and time) and beam display properties of interference (constructive or destructive). High coherence enables precise, focused waveforms; lasers beams are inherently coherent and demonstrate constructive interference "speckles," whereas LEDs are noncoherent sources. The coherence of a laser beam is lost as it enters biologic tissue because of its heterogeneity.
Chromaticity	-	Refers to the luminous wave emitted of similar wavelength and energy; assessed as "band" representing half width of full maximum of a given wavelength; lasers are extremely narrow bands, allowing precise chromophore targeting while sparing adjacent tissue structures.
Polarization	—	Refers to the well-defined direction of electromagnetic field with respect to beam direction (photon); can be linear, circular, or elliptical; lasers are inherently polarized, whereas LEDs and broadband lights are not.

Abbreviations: LED, light-emitting diode; PBM, photobiomodulation.

align well to conventional pharmaceutical approaches to biomedical therapies; therefore, the use of low-dose light therapeutics has been termed *photoceutical*.^{18,19}

Light-tissue interactions are broadly categorized into four processes: absorption, reflection, scattering, and transmission (Figure 2).¹⁰ For the purposes of classifying therapeutic benefits, these interactions can be broadly divided into the pro-

ductive (absorption and scattering, Figures 2A and B) or the nonproductive (reflection and transmission, Figures, 2C and D). The latter processes are exploited for various forms of optical imaging and spectroscopy. These interactions are determined by the physical properties of the light; for example, wavelength, pulsing, total energy and duration, and tissue composition determine the eventual therapeutic efficacy.^{20–23}

ADVANCES IN SKIN & WOUND CARE • VOL. 32 NO. 4 **160** Copyright © 2019 Wolters Kluwer Health, Inc. All rights reserved.

Transmission is the noninteractive passage of a photon through biologic tissue. This is usually seen with high-energy photons and is not used directly for this therapy. Reflection can vary with the angle of the light beam, with the least reflection occurring when treatment beam is perpendicular to tissue.²⁰ Scattering is a result of the heterogeneity of biologic tissues. This process occurs when the incident photon changes its direction of propagation based on differences in refractive indices. Scattering enables the incident light to spread out but progressively reduces penetration, thereby limiting the depth of treatments. Among molecules, dermal collagen has been noted to be predominantly responsible for light scattering. The ability of the skin dermis to act as a turbid matrix results in scattering that approximates an inverse function of wavelength. This implies that shorter (eg, ultraviolet, visible) wavelengths have the greatest scattering, limiting the depth of penetration and increasing the possibility of absorption.

The effectiveness of light-tissue penetration in human skin is predominantly associated with the absorption spectra of three major biologic chromophores: melanin in the epidermis, hemoglobin (oxyhemoglobin and deoxyhemoglobin) in blood within the dermis, and water throughout tissues.^{3,14} Blue (435–500 nm) wavelengths are absorbed by melanin, blood, and porphyrins, whereas red wavelengths (620-750 nm) penetrate deeper because it is subject to less absorption by blood and melanin. These major chromophores have the least impact (absorption) in the nearinfrared (NIR, 750-950 nm) wavelengths, where water becomes more important. If there are no relevant wavelength-specific chromophores in the tissues, the photons pass through the tissue

Figure 2.



LASER-BIOLOGIC TISSUE INTERACTIONS THAT **DETERMINE CLINICAL EFFECTIVENESS**

as total transmission (Figure 2D) without producing any biologic (nonproductive) effects.

Evoked biologic responses are a direct result of the transfer of physical light energy into a biochemical or biophysical change in tissues as the energy progressively attenuates to extinction.^{20,24,25} Therefore, the penetration depth of a given light source is a result of not only its inherent wavelength-dependent photon energy (more for blue-red than NIR) but also the presence of relevant biologic chromophores (also more for blue-red than NIR), with the latter predominating. This essentially implies that effective light penetration is equal to the inverse of the wavelengthspecific tissue absorption coefficient.

Because of this, the two popular PBM wavelengths used are red and NIR, although blue has been more recently explored.^{14,15,26} Red and blue wavelengths are preferred (low penetration, high absorption) for treating superficial tissues, whereas NIR is preferred (low absorption, high penetration) to treat more deepseated tissues (Figure 3).⁸ More recently, research has delineated wavelength-specific responses from specific biologic chromophore absorption versus pure energy transfer responses (Rahman SU, et al, unpublished data, January 2019). Overall, careful examination of light-tissue interactions is critical to enable robust, reproducible clinical treatment parameters.

MOLECULAR MECHANISMS OF PHOTOBIOMODULATION THERAPY

It is increasingly apparent that light-biologic tissue interactions can evoke pathophysiologic (disease) and anatomy-specific responses. The four major biologic responses evoked by PBM include alleviation of pain and inflammation, a modulated immune response, and tissue healing and regeneration.¹² There are three PBM molecular mechanisms that appear to operate within discrete cellular compartments. First, and the most well-studied mechanism, is the absorption of specific wavelengths of light by a key enzyme of the respiratory chain within the mitochondria, cytochrome C oxidase (CCO). Seminal work^{27,28} has demonstrated the action spectrum of CCO runs from the yellow through the red wavelengths (580-700 nm) with discrete peaks around 635 and 730 nm. Absorption of the incident photons by the CCO initiates a photochemical cascade increasing adenosine triphosphate (ATP) and reactive oxygen species (ROS) generation within the electron transport chain. This process involves interaction of ATP synthase with the coenzyme nicotinamide adenine dinucleotide and triggers the combination of inorganic phosphate with adenosine diphosphate to synthesize ATP.

Reactive species are short-lived chemical intermediaries that readily chemically interact with biologic molecules.²⁹ They play critical roles in cell signaling, regulation of cell cycle progression, enzyme activation, and nucleic acid and protein synthesis. Because

161

ADVANCES IN SKIN & WOUND CARE • APRIL 2019 Copyright © 2019 Wolters Kluwer Health, Inc. All rights reserved.

Figure 3.

ABSORPTION SPECTRUM FOR VARIOUS BIOLOGIC CHROMOPHORES THAT AFFECT TISSUE RESPONSES TO SPECIFIC LIGHT WAVELENGTHS⁶¹



The major individual chromophores are noted; although many other biologic elements are known (eg, porphyrins, cryptochromes), their precise spectral characteristics remain to be carefully investigated.

oxygen is a predominant electron acceptor in the electron transport chain, the common ROS include hydrogen peroxide, superoxide, and nitric oxide. In specific cell types (eg, endothelium and cardiac myocytes), small amounts of nitric oxide complexed with CCO are released following light absorption.^{30,31}

A second PBM mechanism focuses on light-modulated cell membrane receptors and transporters such as the opsins, transient receptor potential V1, and aryl hydrocarbon receptor.^{32–34} Absorption of light results in modulation of several key ions such as calcium (Ca^{2+}), protons (H^+), and Na^+/K^+ between the cytosol and extracellular matrix that play key roles in cellular physiology. This specific mechanism is particularly relevant for the analgesic and inflammation modulation responses noted with PBM therapy.

Finally, the third extracellular mechanism has been recently described by the authors,³⁵ involving activation of a multifaceted growth factor, transforming growth factor β (TGF- β), which acts on several cell types in a context-dependent manner. Researchers observed that PBM-induced ROS was sensed by a redoxsensitive methionine that leads to a change in the latent TGF- β 1 conformation, resulting in its activation.

These three mechanisms provide a robust biologic rationale for PBM interventions in specific pathophysiologic contexts. However, several other PBM mechanisms that are relevant to specific anatomical (eg, wounds on limbs vs mucosa) or diseasespecific (eg, venous vs diabetic wounds) pathologies are being actively investigated.

PBM THERAPY FOR WOUND HEALING

There are several well-documented effects of PBM therapy in promoting wound healing.^{36,37} Among the PBM mechanisms outlined previously, extracellular activation of TGF- β 1 appears to be a central wound healing pathway because of its potent effects on a large range of cells in the wound milieu, including hemostasis (platelet-derived TGF- β) and inflammatory cells (macrophage-derived TGF- β) and its prominent role on the extracellular matrix (latent TGF- β -binding protein-associated TGF- β 1 sequestered in the matrix).

Direct effects of PBM on pain and inflammatory mediators such as histamine, serotonin, bradykinin, and prostaglandins have been documented.³⁸ Further, PBM treatments can promote epithelial migration and proliferation, endothelial migration and organization for angiogenesis, inflammatory infiltration, macrophage phagocytoses, immune surveillance, fibroblast matrix synthesis, and wound contraction, among other things.¹² Recent studies have highlighted the efficacy of PBM treatments to promote epithelial cell functions, especially their basal colony-forming units (stem/progenitor cells) that not only can aid re-epithelialization but also promote regeneration of skin appendages such as glands and hair follicles.^{32,39}

There is also now a better understanding of PBM effects on cell lineage-specific responses and maximal dose thresholds that is enabling precise clinical dosing.^{40,41} The PBM dose delivery with a small laser spot size versus multiple diodes or large LED arrays requires careful attention to dosimetry that can have a dramatic effect on healing responses (Figure 4).^{12,42}

Literature Search

On August 17, 2018, authors searched the MeSH terms "PBM" (photobiomodulation therapy) and "LLLT" (low-level laser therapy) and the combination of these with "wound" on PubMed using a filter for studies in humans published in the last 5 years. This exercise was not directed at a comprehensive literature review on PBM and wounds; the 5-year time limit was focused on combining research on current PBM technologies, sound therapeutic rationale, and the current understanding of this treatment modality. Of 218 articles resulting from the search based on these parameters, 11 relevant articles were selected that directly addressed the use of PBM/LLLT applications in chronic wounds. All articles focusing on either surgical debridement or photodynamic therapy that represented higher-powered or thermal effects were excluded.

Search results are discussed here under individual wound types based on increasing complexity of wound etiopathogenesis and treatments with PBM therapy. A specific emphasis is placed on device parameters and clinical delivery protocols, followed by expert author opinion based on both ongoing research and clinical studies.

PBM THERAPY FOR CHRONIC WOUNDS

Chronic wounds are defined as wounds that do not heal for at least 180 days (3 months) and do not proceed through the normal reparative process.⁴³ These wounds usually present with lack of tissue integrity and volume, pain, and persistent inflammation and are often infected.⁴⁴ The initiating injury in these wounds can vary from physical (pressure, burns, or radiation), chemical, electrical, or immunologic injuries that all result in persistent tissue damage. Literature searches for individual clinical presentations of chronic wounds based on inciting injury are described in the following sections.

Burn Wounds

In a recent case series by de Oliveira et al,45 low-dose LED therapy (658 nm red) of varying doses and duration was performed to treat second- and third-degree burn wounds. A total of five cases were included in a double-blind, placebo-controlled (contralateral limb, device switched off) study. The treatment device consisted of a cluster of 12 LEDs with peak power of 40 mW, beam size of 0.13 cm², and an irradiance of 0.31 W/cm². The probe was protected by a translucent film and placed perpendicular to, and in contact with, the skin. The investigators used digital imaging, visual analog scores, and histologic analyses of debrided tissues to assess healing outcomes. All subjects reported less pain and pruritus after PBM treatments, with reduced inflammatory exudate and fibrin with improved re-epithelialization and granulation tissue organization compared with contralateral control sites.45

Expert Opinion. Despite significant heterogeneity among subjects, the investigators reported therapeutic benefits with a single wavelength (red) LED device. Although this device was a 12-LED cluster probe, the authors note treatment limitations of adequately covering large wound surface areas uniformly. Methods to optimize these delivery protocols have been suggested and are being developed, specifically accounting for dosing per unit area (tissue surface irradiance) and treatment time. Interestingly, various doses tested in this study showed some efficacy, and careful calibration in future studies should be investigated.

Venous Ulcers

Vitse et al⁴⁶ recently undertook a prospective, randomized, double-blind, placebo-controlled human clinical study with 24 subjects. The study was designed with two groups; there were no significant differences between subject characteristics in terms of age, gender, or wound parameters. Both groups received standard wound care including debridement of necrotic tissue, hydrating wound dressings, daily application of compression (30 mm Hg), support stockings, and nutrition counseling. The investigators used a 635-nm (red) laser consisting of three diodes at 17.5-mW output power each, with an irradiance of 2.46 mW/cm² for 20-minute treatments, resulting in a total fluence of 2.95 J/cm². Treatments were repeated twice a week for 12 weeks. As controls, the investigators used a sham treatment device indistinguishable from the treatment device to both the subject and treatment-performing investigator. Wounds were assessed at 2 and 12 weeks with arterial blood flow measurements (laser Doppler), pain score (visual analog scale 0-100), record of pain medications taken, and digital photography (laser planimetry). Results included reduced pain in PBM-treated wounds at 4 and 12 weeks. However, there was no significant statistical significance (P > .05) compared with the placebo group.⁴⁶

Expert Opinion. The positive results in this study appear to arise from the reduced sample size and researchers choosing low-responding/nonresponding wounds (ulcer areas 30% following a week of routine wound care). Further, the PBM protocol lacks several key descriptors (delivery details, coverage of wound surfaces), and the dosing appears to be rather inadequate (low irradiance-fluence, lack of adequate repetitions). Optimization of these parameters in future studies, along with routine venous wound management approaches, could promote PBM therapy as a potent adjuvant in venous ulcer care.47 Further, specific emphasis on dose and wavelength from laboratory studies demonstrating the ability of PBM treatments to modulate endothelial biology could be very informative.⁴⁸

Pressure Ulcers

In a well-designed study, Taradaj et al⁴⁹ performed a singleblind, randomized clinical study to assess the effects of three common wavelengths to treat pressure ulcers. Their study recruited 71 subjects and placed them in four groups: (1) placebo (routine standard of wound care and laser treatments with device switched off), (2) 658 nm (red laser), (3) 808 nm (NIR laser), and (4) 940 nm (NIR laser). All groups received routine care for their pressure wounds including daily wound irrigations with a 0.9% physiologic saline solution and 1% hydrophilic silver sulfadiazine cream along with adapted footwear, self-care, and the prevention of disabilities.⁵⁰ Wound areas were assessed with planimetry using an NIR camera. The lasers were used at a power output of 50 mW with a spot size of 0.1 cm² for an effective fluence of 4 J/cm² at the tissue surface. Treatments were performed with a scanner 50 cm from the wound surface with a movement frequency at 20 Hz along the ordinate axis and 0.5 Hz along the abscissa axis. All groups were homogenous in all participant characteristics. Their results noted that the 658-nm laser treatment was most effective (70% closure, P < .05) at promoting wound closure. In contrast, the 808- and 940-nm laser treatments (31% and 30% closure, respectively) did not appear to improve healing rates significantly compared with the placebo group (28% closure).49

Figure 4.

THE DIFFERENCES BETWEEN LASER (A), SINGLE-LED (B), OR MULTIPLE-LED ARRAY (C) ILLUMINATION OF BIOLOGIC TISSUES⁶²



While the laser beam is shown as a coherent and collimated beam that loses these properties when it enters the tissues, the LEDs have a noncoherent, divergent beam that inherently lacks these characteristics. Nonetheless, both light sources have been shown to generate therapeutic PBM responses. The LED array emphasizes the increased beam intensity (which partly overcomes LED lack of tissue penetration) and larger area of tissue coverage that significantly improves its utility compared with the more efficient laser PBM device.

Expert Opinion. The use of a scanning mode with the limited spot size of lasers requires consistent technique to effectively deliver care. This study did a good job of outlining the methods used and represents an effective starting point to develop a more rigorous protocol.⁴⁶ Unfortunately, a lack of attention to several PBM treatment parameters has resulted in inconsistent outcomes that are of questionable utility in clinical practice guidelines.⁵⁰ Attention to both single treatment dosing (time, power density, and wound coverage pattern) and repetitions (per week and total number of treatments) can further improve future clinical outcomes. Recent evidences have observed specific (but not all) biologic responses appear to respond therapeutically when lower power (energy per photon) of NIR wavelength lasers is compensated with higher visible wavelengths (red photons). Researchers should take care to avoid the thermal effects of NIR and direct attention to their increased tissue penetration. The use of other wavelengths (eg, blue), as mentioned by Taradaj et al,⁴⁹ could be helpful to manage pain and infections.

Diabetic Foot Ulcers

In a randomized controlled study by Feitosa et al⁵¹ in 2015, 16 subjects with uncontrolled diabetes and diabetic foot ulcers were divided into two groups. Both groups received routine wound care treatments, whereas one group was also treated with a 30-mW laser at 632.8 nm for a fluence of 4 J/cm². Each treatment lasted 80 seconds and was repeated three times a week for 4 weeks. Wounds were assessed by digital imaging, and a visual analog pain score was used. Their results demonstrated a statistically significant (P < .05) reduction in wound size and pain scores in PBM-treated wounds at 12 and 30 days compared with controls.⁵¹

In a study by Mathur et al,⁵² 30 controlled subjects with diabetes (fasting blood sugar 200 mg/dL) with at least one diabetic foot ulcer persisting for 6 weeks were treated with a 660-nm laser at 50 mW/cm² for 60 seconds daily for 15 days. The investigators assessed wounds with laser Doppler imaging, anterior brachial index, and digital wound imaging. They noted a significant improvement in PBM-treated wounds (30%–50%, P < .05) in a majority of subjects compared with minimal changes in the control group.⁵²

Figure 5.

THE MODES OF LIGHT SOURCE OPERATIONS CAN BE CONTINUOUS (A) OR PULSED (B, C).



The pulsing characteristics are defined by the pulse duration (width), reported in seconds; pulse intervals, reported in seconds; and pulse frequency (number of pulses per second, in Hz). In more sophisticated pulsing sequences, especially in ultrafast pulses (nanoseconds, picoseconds, or fentoseconds), a series of pulses termed the pulse train are followed by a pulse gap. This is usually reported as time (eg, picoseconds). The duty cycle refers to the time the beam is on during entire treatment. For a continuous wave, this is 100%, although it can vary significantly with pulsing regimens as noted here.

ADVANCES IN SKIN & WOUND CARE • VOL. 32 NO. 4 Copyright © 2019 Wolters Kluwer Health, Inc. All rights reserved.

In another study, Carvalho et al⁵³ examined the utility of combining essential fatty acids (EFAs) as a wound dressing along with PBM treatments. They recruited 32 subjects with controlled diabetes (fasting blood sugar 150–350 mg/dL) and placed them into one of four groups: (1) control (no interventions), (2) EFAs alone, (3) PBM alone, or (4) PBM-EFA treatments. They obtained similar laser Doppler imaging, anterior brachial indices, visual analog scale pain scores, and digital wound imaging where they observed a statistically significant reduction (P < .05) in pain and wound area in the PBM and PBM-EFA groups compared with controls and EFAs alone.⁵³

Finally, Ruh et al⁵⁴ investigated the potential benefits of PBM therapy in a case series with eight subjects with diabetes who were immobilized for at least 12 months and presented with pressure ulcers (grades II-IV). They performed PBM treatments with a 100-mW laser at 660 nm for 12 seconds per point, with each point 2 cm apart for total fluence of 2 J/cm². Treatments were performed daily for 12 days; wounds were analyzed clinically, and tissues were collected for mRNA analyses for cytokine profiles. Their results demonstrated significant improvements in wound areas following PBM treatments that correlated with increased prohealing factors including vascular endothelial growth factor (angiogenesis) and TGF-B (matrix and epithelial closure), whereas TNF- α (inflammation) levels were reduced. These results indicate that the clinical improvements noted with PBM treatments appear to correlate with tissue factors representing improved healing responses.54

Expert Opinion. The two major points highlighted in these studies are that attention to the clinical protocol should be based on underlying pathology (metabolic, inflammation, or neurological) as well as optimized delivery regimens.³⁸ Direct assessment of "tissue surface irradiance" provides a more reliable delivery parameter than average power of the unit, irradiance at probe output, or distance-to-target measurements.⁵⁵ Changes in molecular markers offer further objective assessment of wound healing responses, strengthening the evidence for any therapeutic benefits.56

CONCLUSIONS

This brief literature review indicates the benefits of PBM therapy for various types of wounds. There are several studies on the therapeutic efficacy of PBM therapy for oral mucositis, ultraviolet skin damage, and radiation dermatitis that were outside the scope of this review.^{42,57–59}

A major limitation noted in this review was the wide range of PBM clinical protocols among studies with respect to wavelength, dose, and delivery that prevent a rigorous consensus. Understanding the photobiologic mechanisms of PBM using correct wavelengths is critical for optimal clinical light treatment parameters for desired therapeutic medical and biological outcomes. Nonetheless, within the limits of the presented studies, the clinical benefits noted could serve as templates for development of more rigorously designed clinical studies to evaluate PBM therapy in wound healing.

Attention to treatment costs such as purchase and maintenance of PBM equipment should be balanced with potential improved clinical efficacy and the benefits of lowered care costs. Further, the access to wound care afforded by this nonpharmacologic intervention could also be specifically evaluated. Among various avenues for further development of PBM technologies, a photoceutical approach to validating pharmacodynamics, pharmacokinetics, and clinical safety and efficacy seems most attractive. Targeting robust, objective outcomes, especially molecular biomarkers, will enable PBM therapy to become a potent modality for wound management in the future.

PRACTICE PEARLS

• Promotion of wound healing was among the first observed benefits of low-dose light treatments (photobiomodulation therapy). There is now growing evidence from well-designed controlled human studies for this specific application.

- Discrete wavelengths and light dose can be used to promote specific phases of wound healing.
- There are now well-understood molecular mechanisms mediating the use of low-dose light therapies.
- · Light treatment parameters can be tailored to specifically account for both anatomical and pathophysiologic responses, including device settings and treatment delivery approaches.

• As with most current therapies, the biologic target and disease etiology must be addressed for maximal reproducible therapeutic benefit from these treatments.

REFERENCES

- 1. Grzybowski A, Sak J, Pawlikowski J. A brief report on the history of phototherapy. Clin Dermatol 2016;34(5):532-7.
- 2. Grzybowski A, Pietrzak K. From patient to discoverer-Niels Ryberg Finsen (1860-1904)-the founder of phototherapy in dermatology. Clin Dermatol 2012;30(4):451-5.
- 3. Patil UA, Dhami LD. Overview of lasers. Indian J Plast Surg 2008;41(Suppl):S101-13.
- 4. Rocchietta S. Notes for a history of physiotherapy. Niels Ryberg Finsen (1860-1904), pioneer of modern phototherapy, on the centenary of his birth [in Italian]. Minerva Med 1960;51:966-8.
- 5. Mester E, Spiry T, Szende B, Tota JG. Effect of laser rays on wound healing. Am J Surg 1971;122(4):532-5.
- 6. Mester E, Szende B, Gartner P. The effect of laser beams on the growth of hair in mice [in German]. Radiobiol Radiother (Berl) 1968;9(5):621-6.
- 7. Mester E, Szende B, Spiry T, Scher A. Stimulation of wound healing by laser rays. Acta Chir Acad Sci Hung 1972;13(3):315-24.
- 8. Anders JJ, Lanzafame RJ, Arany PR. Low-level light/laser therapy versus photobiomodulation therapy. Photomed Laser Surg 2015;33(4):183-4.
- 9. Sousa RG, Batista Kde N. Laser therapy in wound healing associated with diabetes mellitus-review. An Bras Dermatol 2016;91(4):489-93.

165

- Zhu Y, Xu G, Yuan J, et al. Light emitting diodes based photoacoustic imaging and potential clinical applications. Sci Rep 2018;8(1):9885.
- 11. Nelson JS. Lasers: state of the art in dermatology. Dermatol Clin 1993;11(1):15-26.
- Arany PR. Craniofacial wound healing with photobiomodulation therapy: new insights and current challenges. J Dent Res 2016;95(9):977-84.
- Avci P., et al., Low-level laser (light) therapy (LLLT) in skin: stimulating, healing, restoring. Semin Cutan Med Surg 2013; 32(1):41-52.
- Chung H, Dai T, Sharma SK, Huang YY, Carroll JD, Hamblin MR. The nuts and bolts of low-level laser (light) therapy. Ann Biomed Eng 2012;40(2):516-33.
- Heiskanen V, Hamblin MR. Photobiomodulation: lasers vs. light emitting diodes? Photochem Photobiol Sci 2018;17(8):1003-17.
- Reinisch L. Laser physics and tissue interactions. Otolaryngol Clin North Am 1996;29(6): 893-914.
- Parrish JA. Dohi memorial lecture. Laser medicine and laser dermatology. J Dermatol 1990;17(10):587-94.
- Lin F, Josephs SF, Alexandrescu DT, et al. Lasers, stem cells, and COPD. J Transl Med 2010;8:16.
- Karu T. Is it time to consider photobiomodulation as a drug equivalent? Photomed Laser Surg 2013;31(5):189-91.
- 20. Carroll L, Humphreys TR. LASER-tissue interactions. Clin Dermatol 2006;24(1):2-7.
- Jacques SL. Ratio of entropy to enthalpy in thermal transitions in biological tissues. J Biomed Opt 2006;11(4):041108.
- Jacques SL. Path integral description of light transport in tissue. Ann N Y Acad Sci 1998;838:1-13.
- Wang L, Liu D, He N, Jacques SL, Thomsen SL. Biological laser action. Appl Opt 1996;35(10):1775-9.
- Wanner M, Sakamoto FH, Avram MM, Anderson RR. Immediate skin responses to laser and light treatments: warning endpoints: how to avoid side effects. J Am Acad Dermatol 2016;74(5):807-19.
- Anderson RR. Lasers for dermatology and skin biology. J Invest Dermatol 2013;133(E1): E21-3.
- Fornaini C, Merigo E, Rocca JP, et al. 450 nm blue laser and oral surgery: preliminary ex vivo study. J Contemp Dent Pract 2016;17(10):795-800.
- Karu TI1, Pyatibrat LV, Kolyakov SF, Afanasyeva NI. Absorption measurements of a cell monolayer relevant to phototherapy: reduction of cytochrome C oxidase under near IR radiation. J Photochem Photobiol B 2005;81(2):98-106.
- Wong-Riley MT, Liang HL, Eells JT, et al. Photobiomodulation directly benefits primary neurons functionally inactivated by toxins: role of cytochrome C oxidase. J Biol Chem 2005;280(6):4761-71.
- Halliwell B, Gutteridge JM. The definition and measurement of antioxidants in biological systems. Free Radic Biol Med 1995;18(1):125-6.
- Keszler A, Lindemer B, Hogg N, Weihrauch D, Lohr NL. Wavelength-dependence of vasodilation and NO release from S-nitrosothiols and dinitrosyl iron complexes by far red/near infrared light. Arch Biochem Biophys 2018;649:47-52.
- Moriyama Y, Moriyama EH, Blackmore K, Akens MK, Lilge L. In vivo study of the inflammatory modulating effects of low-level laser therapy on iNOS expression using bioluminescence imaging. Photochem Photobiol 2005;81(6):1351-5.
- Buscone S, Mardaryev AN, Raafs B, et al. A new path in defining light parameters for hair growth: discovery and modulation of photoreceptors in human hair follicle. Lasers Surg Med 2017;49(7):705-18.
- Becker A, Klapczynski A, Kuch N, et al. Gene expression profiling reveals aryl hydrocarbon receptor as a possible target for photobiomodulation when using blue light. Sci Rep 2016;6:33847.
- Wang Y, Huang YY, Wang Y, Lyu P, Hamblin MR. Photobiomodulation of human adiposederived stem cells using 810nm and 980nm lasers operates via different mechanisms of action. Biochim Biophys Acta Gen Subj 2017;1861(2):441-9.
- Arany PR, Cho A, Hunt TD, et al. Photoactivation of endogenous latent transforming growth factor-beta1 directs dental stem cell differentiation for regeneration. Sci Transl Med 2014;6(238):238ra69.
- Woodruff LD, Bounkeo JM, Brannon WM, et al. The efficacy of laser therapy in wound repair: a meta-analysis of the literature. Photomed Laser Surg 2004;22(3):241-7.
- 37. Peplow PV, Baxter GD. Gene expression and release of growth factors during delayed wound healing: a review of studies in diabetic animals and possible combined laser

phototherapy and growth factor treatment to enhance healing. Photomed Laser Surg 2012;30(11):617-36.

- Rohringer S, Holnthoner W, Chaudary S, et al. The impact of wavelengths of LED lighttherapy on endothelial cells. Sci Rep 2017;7(1):10700.
- Khan I, Arany PR. Photobiomodulation therapy promotes expansion of epithelial colony forming units. Photomed Laser Surg 2016;34(11):550-5.
- Khan I, Tang E, Arany P. Molecular pathway of near-infrared laser phototoxicity involves ATF-4 orchestrated ER stress. Sci Rep 2015;5:10581.
- Engel KW, Khan I, Arany PR. Cell lineage responses to photobiomodulation therapy. J Biophotonics 2016;9(11-12):1148-56.
- Rahman SU, Mosca RC, Govindool Reddy S, et al. Learning from clinical phenotypes: low-dose biophotonics therapies in oral diseases. Oral Dis 2018;24(1-2):261-76.
- Frykberg RG, Banks J. Challenges in the treatment of chronic wounds. Adv Wound Care (New Rochelle) 2015;4(9):560-82.
- Schultz GS, Sibbald RG, Falanga V, et al. Wound bed preparation: a systematic approach to wound management. Wound Repair Regen 2003;11 Suppl 1:S1-28.
- De Oliveira RA, Boson LLB, Portela SMM, Filho ALMM, de Oliveira Santiago D. Lowintensity LED therapy (658 nm) on burn healing: a series of cases. Lasers Med Sci 2018;33(4):729-35.
- Vitse J, Bekara F, Byun S, Herlin C, Teot L. A double-blind, placebo-controlled randomized evaluation of the effect of low-level laser therapy on venous leg ulcers. Int J Low Extrem Wounds 2017;16(1):29-35.
- Brem H, Kirsner RS, Falanga V. Protocol for the successful treatment of venous ulcers. Am J Surg 2004;188(1A Suppl):1-8.
- 48. Amini A, Pouriran R, Abdollahifar MA, et al. Stereological and molecular studies on the combined effects of photobiomodulation and human bone marrow mesenchymal stem cell conditioned medium on wound healing in diabetic rats. J Photochem Photobiol B 2018;182:42-51.
- Taradaj J, Halski T, Kucharzewski M, Urbanek T, Halska U, Kucio C. Effect of laser irradiation at different wavelengths (940, 808, and 658 nm) on pressure ulcer healing: results from a clinical study. Evid Based Complement Alternat Med 2013;2013:960240.
- National Pressure Ulcer Advisory Panel. 2014 Prevention and Treatment of Pressure Ulcers: Clinical Practice Guideline. 2014. www.npuap.org/resources/educational-andclinical-resources/prevention-and-treatment-of-pressure-ulcers-clinical-practiceguideline. Last accessed January 10, 2019.
- Feitosa MC, Carvalho AF, Feitosa VC, Coelho IM, Oliveira RA, Arisawa EÂ. Effects of the low-level laser therapy (LLLT) in the process of healing diabetic foot ulcers. Acta Cir Bras 2015;30(12):852-7.
- Mathur RK, Sahu K, Saraf S, Patheja P, Khan F, Gupta PK. Low-level laser therapy as an adjunct to conventional therapy in the treatment of diabetic foot ulcers. Lasers Med Sci 2017;32(2):275-82.
- Carvalho AF, Feitosa MC, Coelho NP, et al. Low-level laser therapy and *Calendula* officinalis in repairing diabetic foot ulcers. Rev Esc Enferm USP 2016;50(4):628-34.
- Ruh AC, Frigo L, Cavalcanti MFXB, et al. Laser photobiomodulation in pressure ulcer healing of human diabetic patients: gene expression analysis of inflammatory biochemical markers. Lasers Med Sci 2018;33(1):165-71.
- Khan I, Arany PR. Dosimetry for photobiomodulation therapy: response to Sommers et al. Ann Transl Med 2016;4(10):208.
- Brem H1, Stojadinovic O, Diegelmann RF, et al. Molecular markers in patients with chronic wounds to guide surgical debridement. Mol Med 2007;13(1-2):30-9.
- Leventhal J, Young MR. Radiation dermatitis: recognition, prevention, and management. Oncology (Williston Park) 2017;31(12):885-7, 894-9.
- Censabella S, Claes S, Robijns J, Bulens P, Mebis J. Photobiomodulation for the management of radiation dermatitis: the DERMIS trial, a pilot study of MLS() laser therapy in breast cancer patients. Support Care Cancer 2016;24(9):3925-33.
- Gautam AP, Fernandes DJ, Vidyasagar MS, Maiya AG, Guddattu V. Low level laser therapy against radiation induced oral mucositis in elderly head and neck cancer patients-a randomized placebo controlled trial. J Photochem Photobiol B 2015;144:51-6.
- 60. Weber MJ. Handbook of Laser Wavelengths. 1st ed. Boca Raton, FL: CRC Press; 1999.
- Olivi G, Olivi M. Laser-hard tissue interaction. In: Olivi G, Olivi M, eds. Lasers in Restorative Dentistry. 1st ed. Berlin, Germany: Springer; 2015.
- Calderhead RG, Tanaka Y. Photobiological basics and clinical indications of phototherapy for skin rejuvenation. In: Tanaka Y, ed. Photomedicine: Advances in Clinical Practice. 1st ed. London, UK: IntechOpen; 2017:216-56.

For more than 144 additional continuing education articles related to Skin and Wound Care topics, go to NursingCenter.com/CE.

CONNECTION

CONTINUING MEDICAL EDUCATION INFORMATION FOR PHYSICIANS

Lippincott Continuing Medical Education Institute, Inc., is accredited by the Accreditation Council for Continuing Medical Education to provide continuing medical education for physicians.

Lippincott Continuing Medical Education Institute, Inc., designates this journal-based CME activity for a maximum of 1 AMA PRA Category 1 CreditTM. Physicians should claim only the credit commensurate with the extent of their participation in the activity.

PROVIDER ACCREDITATION INFORMATION FOR NURSES

Lippincott Professional Development will award 1.5 contact hours for this continuing nursing education activity

LPD is accredited as a provider of continuing nursing education by the American Nurses Credentialing Center's Commission on Accreditation.

This activity is also provider approved by the California Board of Registered Nursing, Provider Number CEP 11749 for 1.5 contact hours. LWW is also an approved provider by the District of Columbia, Georgia, and Florida CE Broker #50-1223.

OTHER HEALTH PROFESSIONALS

This activity provides ANCC credit for nurses and AMA PRA Category 1 CreditTM for MDs and

DOs only. All other healthcare professionals participating in this activity will receive a certificate of participation that may be useful to your individual profession's CE requirements

CONTINUING EDUCATION INSTRUCTIONS

- Read the article beginning on page 157. For nurses who wish to take the test for CNE contact hours, visit http://nursing.ceconnection.com. For physicians who wish to take the test for CME credit, visit http://cme.lww.com. Under the Journal option, select Advances in Skin and Wound Care and click on the title of the CE activity
- You will need to register your personal CE Planner account before taking online tests. Your planner will keep track of all your Lippincott Professional Development online CE activities for you,
- There is only one correct answer for each question. A passing score for this test is 13 correct answers. If you pass, you can print your certificate of earned contact hours or credit and access the answer key. Nurses who fail have the option of taking the test again at no additional cost. Only the first entry sent by physicians will be accepted for credit.

Registration Deadline: March 31, 2021 (physicians); March 5, 2021 (nurses).

PAYMENT

• The registration fee for this CE activity is \$17.95 for nurses; \$22.00 for physicians

CALL FOR PAPERS Advances in SKIN **VOUNDCARE®** Advances in Skin & Wound Care is soliciting manuscripts on a variety of topics, including original investigations,

clinical concepts, and literature reviews. Consult the journal's

Web site, www.woundcarejournal.com, for the

Author Guidelines.

ADVANCES IN SKIN & WOUND CARE · APRIL 2019 Copyright © 2019 Wolters Kluwer Health, Inc. All rights reserved.